CUI INC
a bel group

CE

Declaration of Conformity

Manufacturer:
CUI Inc.
15575 SW Sequoia Parkway, Suite 100
Portland, Oregon 97224

For the following equipment:

AC-DC Internal Power Supply
CUI Series: VMS-60
Models: See next page

This declaration of conformity is issued under the sole responsibility of the manufacturer.
The object of the declaration described above is in conformity with the relevant Union harmonization legislations and their amendments:

General Product Safety Regulation (EU)/2023/988

EMC Directive 2014/30/EU

RoHS Directive 2011/65/EU and (EU) 2015/863

References to the relevant harmonized standards used, including the date of the standard, or references to the other technical specifications, including the date of the specification, in relation to which conformity is declared:

Health & Safety	EN 60601-1:2006+A1:2013+A12:2014+A2:2021, IEC 60601-1:2005+A1:2012+A2:2020 (equivalent)
EMC	EN 60601-1-2:2015+A1:2021, IEC 60601-1-2:2014+A1:2020 (equivalent); EN 55011:2016+A11:2020+A2:2021; EN IEC 61000-3-2:2019+A1:2021; EN 61000-3-3:2013+A1:2019+A2:2021
RoHS	EN IEC 63000:2018

Note: These component level power supplies are intended exclusively for inclusion within other equipment. Protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

	Shenzhen, China	**03/07/2025**
(manufacturer)	(place)	(date)

Link Lu
Product Compliance Specialist

	Portland, Oregon, USA	**03/07/2025**
(manufacturer)	(place)	(date)

Editha Vergara
Global Director, Safety, Environmental

MODEL LIST

VMS-60-XX (where XX = 5, 12, 15, 24, 48 denote output voltage)

Model	Input voltage (Vac)	Frequency (Hz)	Input current (A)	Output voltage (Vdc)
VMS-60-5	100-240	47-63	1.6-0.8	5
VMS-60-12	100-240	47-63	1.6-0.8	12
VMS-60-15	100-240	47-63	1.6-0.8	15
VMS-60-24	100-240	47-63	1.6-0.8	24
VMS-60-48	100-240	47-63	1.6-0.8	48
Note: Model name maybe followed by additional characters as described on below model naming configuration.				

Model Naming Configuration

VMS	**-**	**60**	**-**	**XX**
	I		-	II

I – Base Number: VMS-60

II – Output Voltage: 5 = 5 V; 12 = 12 V; 15 = 15 V; 24 = 24 V; 48 = 48 V

REVISION HISTORY

rev.	description	date
1.0	initial release	06/14/24
1.01	Change MDR Regulation 2017/745/EU to General Product Safety Regulation (EU)/2023/988. Update EN 55011 to amendment 2, EN 61000-3-3 to amendment 2	03/07/25

The revision history provided is for informational purposes only and is believed to be accurate.