CERTIFICATE

No. P15220106

Order No. 290254

Product	AC-DC switching power supply
Applicant	Bel Fuse Inc. 206 Van Vorst St. Jersey City, NJ 07302 USA
Manufacturer	Bel Fuse Inc. 206 Van Vorst St. Jersey City, NJ 07302 USA
Factory	BPS Asia Pacific Electronics (Shenzhen) Co., Ltd. Building# 6, Nanming Road, Gongming Town Huahong Xintong Industrial Park, Guangming District Shenzhen 518108 China ☐ See next page(s)
Ratings	Input: 8.0-4.0 A, 100-250 Vac, 50/60 Hz or 7.0-3.5 A, 120-240 Vdc
Trade mark	bel POWER SOLUTIONS & PROTECTION a bel group
Model / Type Ref.	MBC600-1012G; MBC600-1024G; MBC600-1048G
Principal characteristics	Cl. I. DC outputs: V1: 50 A, 12 V for model MBC600-1012G or 25 A, 24 V for model MBC600-1024G or 12.5 A, 48 V for model MBC600-1048G / Vsb: 1.0 A, 5 V / Vfan: 1.0 A, 12 V. The risk management requirements of the standard were not addressed. ☐ See next page(s)
A sample of the product was tested and found to be in conformity with	MED EN 60601-1:2006;A1
Validity	This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.
Additional information	☐ See next page(s) The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and

Date of issue 12-08-2015

Juan Z. Kleppenes

Juan Z. Kleppenes
Certification Department

Nemko AS
Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532

Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

Additional model(s)

☐ See next page(s)

Date of issue 12-08-2015

Juan Z. Kleppenes

Certification Department

Nemko AS
Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532